



التاريخ : 03/06/2018

إعادة طرح عطاء رقم – RFQ-H-10958

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين
الموضوع : أجهزة ومستهلكات طبية ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (350) شيكل إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة – الطابق الأول .
3. كفالة دخول عطاء بقيمة (1000) ألف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم التأمين لا ينظر في عرض سعركم.
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر .
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
7. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
8. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر وتحملكم فروق الأسعار.
9. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
10. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
11. التوريد جزئي ووفقاً لحاجة المستشفى .
12. أن لا تقل صلاحية استخدام المواد عند توريدها عن سنتين .
13. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح أي منهما العرض الفني والعرض المالي.
14. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
15. ضرورة الرد على العطاء سلباً او ايجاباً .
16. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
17. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتزليل .
18. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
19. اخر موعد لتقديم عرض السعر هو يوم الخميس الموافق 2018-06-14 الساعة الثالثة عصراً 03:00 PM .
20. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (97092389687+) داخلي (6400) او من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

23680/AM

إياد المكاي



إعادة طرح عطاء رقم – RFQ-H-10958

No.	Item	QTY	Unit	Supplier & Code No.	delivery	Unit Price (NIS)	Total Price (NIS)
1	Fractional CO2 Laser : Dot/ Radiofrequency laser type: Co2 RF wavelength : 10.6 μm Emission time : 0.01 to 0.9 sec Emission mode: CW, SP, DP, HP, UP Aiming beam : laser diode at 635 nm control panel : wide LCD color touch screen 10" scanning shapes : DOT , line , Triangle , parallelogram , exagon ,square Scanning modes: normal,interlaced and smart scan RF power : 1 to 50w suitable UPS should include in the main price smoke evacuator with tubing and filter Smart cryo system with connection operation and service training should include in the main price	1	Pcs				



No.	Item	QTY	Unit	Supplier & Code No.	delivery	Unit Price (NIS)	Total Price (NIS)
2	Square Epil Laser Source: Alexandrite and Nd:YAG laser heads Wavelength : 755 + 1064 nm Max Energy : 135 J Pulse Duration: 3-100 ms Pulse Repetition Rate: 3 Hz Max Round Spot Size Ø: 30 mm Max Square Spot Size: 27x27 mm Skin Cooling System: Cryo air and contact Standard handpiece with the following round spot sizes : 2 - 3 - 5 - 10 - 14 18 - 20 Footswitch should available optional accessories : should be priced separately Square Spot sizes Square sport sizes up to 27x27mm Round spot size up to 30 mm	1	Pcs				
3	Handpiece Smoke Evacuator With tubing and filters UV Light and derma marker operation and service training should include in the main price	1	Pcs				

ln



No.	Item	QTY	Unit	Supplier & Code No.	delivery	Unit Price (NIS)	Total Price (NIS)
4	Synchro vasQ (pulse dye laser) laser type : dye laser wavelength : 595nm spot dimension : 5,7,10 and 12 mm optional spot: spacer with compression lens for pigmented 3mm , 3x10 mm and 9x9 mm pulse duration : 0.3 to 40 ms repetition rate : 0.2 to 1 Hz aiming beam : 4mW at 532 nm emission control : foot switch or finger switch (specify) database : programmable control panel : LCD TFT 10" or better Pulsed dye lamp should be including in the main price smoke evacuator with tubing and filter and cooling system operation and service training should include in the main price .	1	Pcs				
5	Synchro QS or MDK III laser type : QS ND: YAG& Pulsed wavelength : 1064 , 532 nm Pulse width: 6 ns in QS mode , 2-25 MS ms in pulsed mode Max Energy: 1000 mJ (QS ND: YAG) , 3 mJ (Pulsed) Max Fluence: 25-95 (J/CM) Repetition :20 Hz Spot sizes: 5x5, 4x4, 3x32x2 MM Spot size: 2x2-6x6 mm or up to 9X9 Beam Delivery: Articulated Arm Smoke evacuator with tubings and filters. Cooling system operation and service training should include in the main price .	1	Pcs				

em



No.	Item	QTY	Unit	Supplier & Code No.	delivery	Unit Price (NIS)	Total Price (NIS)
6	Excilite -Micro type of source : Excimer XeCl wavelength : 308 nm Power Density : 50 mW/cm ² Emission : continues 1-90 s treatment surfaces : Rectangular , circular control panel : LCD Touch screen	1	Pcs				
7	Dermapen Controlled delivery The ability to control the delivery any type of water based Meso Solution in incrementals of 0.01ml per/ second The ability to control the depth of the injection also Precision To deliver both DP Dermaceuticals protocol Meso-Serums as well as fluids such a PRP, controlled directly to the treatment area around the needles. Flexibility Meso-Serums can be supported in syringe sizes from 1cc to 50cc syringes. Reliable The Dermapen infuse Dermapen inject can offer both the dermapen and the carboxytheraoy in the same machine delivers the TattOff Activator Serum in a continuous reliable flow	1	Pcs				



No.	Item	QTY	Unit	Supplier & Code No.	delivery	Unit Price (NIS)	Total Price (NIS)
	to enable unparalleled fluid delivery needle - tips should include in the main price Also need to determine the price of the needles separately (disposable needles)						
8	carboxy therapy equipment -Convenient: Mini pen No : not the mini pen is the needed device. Ability to control pressure of infusion and quantity infused. Tubing and CO2 container -Safe: No-side effect: using super pure C2P injection -should Quick and Effective - No down time	1	Pcs				
9	PRP kit High Concentration High Platelet Recovery Easy to use Fits Standard Centrifuge Sterile device Clear Separation of RBC and PPP	100	Pcs				
	Total Price (NIS)						

- يرجى من حضرتكم تحديد مدة التوريد ومدة الكفالة لكل جهاز.

23680/AM

ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء



مستشفى النجاح الوطني الجامعي

إعلان إعادة طرح عطاء

تدعو مستشفى النجاح الوطني الجامعي الشركات المختصة والراغبة بالمشاركة في العطاء المذكور في الجدول:

موعد الإغلاق	#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	موعد طرح العطاء	التاريخ	الساعة
PM 03:00	1	<u>RFQ-H-10958</u>	أجهزة ومستهلكات طبية	<u>350</u> شيكل غير مستردة	2018/06/03	2018/06/14	

وحسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الإعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 350 شيكل غير مستردة .
4. كفالة دخول عطاء بقيمة (1000) ألف دولار أمريكي .

على الراغبين بالإشتراك في العطاء المذكور طلبه بواسطة ارسال رسالة على الإيميل التالي tender3@najah.edu او مراجعة دائرة اللوازم والمشتريات مكتب رقم ١6400 في مستشفى النجاح الوطني الجامعي مبنى الإدارة ، للحصول على نسخة من وثائق العطاء.

للاستفسار يرجى الاتصال على:

دائرة اللوازم والمشتريات في مستشفى النجاح الوطني الجامعي – البريد الالكتروني tender3@najah.edu هاتف 970-09-2389687 - فاكس: 970-9-2389685 .