

التاريخ : 04-10-2021

طرح عطاء رقم – RFQ-H-11429

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : **Complete Flexible Endoscopy Endoscopy Station** ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (500) شيكل اسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة - الطابق الأول .
3. كفالة دخول عطاء بقيمة (2000) ألفا دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم التأمين لاينظر في عرض سعركم.
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر تستبدل بكفالة صيانة بقيمة 5% من قيمة الإحالة .
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
7. أن تكون الأجهزة المعروضة جديدة وليست مجددة .
8. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
9. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر وتحملكم فروق الأسعار.
10. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.
13. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
14. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الاسعار .
15. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
16. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
17. آخر موعد لتقديم عرض السعر هو يوم الخميس الموافق 2021-10-14 الساعة الثالثة عصراً 03:00 PM .

18. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي

طرح عطاء رقم – RFQ-H-11429

No.	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	U.P (NIS)	T.P (NIS)
1	G-A-MDD012	Complete Flexible Endoscopy Station	2	Pcs				
-	Total Price (NIS)							

#	Complete Flexible Endoscopy Station, QTY 2	Compliance	Deviation	Unit price (NIS)	Total Price (NIS)
	Good brand name with FDA or CE (Europe, USA , Japan)				
# 1	Mobile endoscopy Cart, Qty: 2				
	Manufacturer:				
	Source and origin:				
	Model:				
	Mobile cart with 4 shelves , 1 drawer, Key board tray, Monitor arm, Scopes holder with front breaks and lateral bumpers. With main switch and multiple sockets block (Qty.: 6)				
	Complete with hanger for irrigation bottle				
	A 220V, 50Hz electrical source				
	Protections against over-voltage and over-current line conditions.				
# 2	HD Monitor, Qty: 4				
	Manufacturer:				
	Source and origin:				
	Model:				
	One monitor for main station and other for wall				
	LCD monitor for medical applications with Ultra High Definition picture quality, size not less than 25 inch for main station and 40 inch for wall monitor				

	The monitor should have input-output for analog and digital signals and suitable for HD signals				
	A 220V, 50Hz electrical source				
	Protections against over-voltage and over-current line conditions.				
# 3	Endoscopy system, compact design or individual:				
# 3.1	HD Video Processor, Qty: 2				
	Manufacturer:				
	Source and origin:				
	Model:				
	Digital signal video processor for color high definition and high resolution endoscopy images				
	Please specify the observation modes available				
	Digital outputs: DVI, HD SDI, HDMI and VGA				
	Auto iris modes can be selected using the "iris mode" switch on the front panel.				
	Automatic white balance, color adjustment and image enhancement.				
	Digital zoom				
	Fog free function				
	Image sizing and freezing modes with switching the enhancement mode.				
	Advance registration of patient data, and doctor database				
	Internal memory				
	System should be complete with all cabling and connectors accessories with endoscopes, light source and monitor				
	A 220V, 50Hz electrical source				
	Protections against over-voltage and over-current line conditions.				
# 3.2	Light source, Qty: 2				
	Manufacturer:				
	Source and origin:				
	Model:				
	Separate or built in controller				

	Multiple LED technology, Please specify the number of LED				
	Specify the light life time expected				
	Intensity control 0 up to 100 %				
	Microprocessor controlled				
	Automatic and manual variable light control				
	If the light source contain emergency light, please specify it				
	Simple and safe scopes attachment, one step connector				
	A 220V, 50Hz electrical source				
	Protections against over-voltage and over-current line conditions.				
# 4	HD Gastroscope, Qty: 4				
	Manufacturer:				
	Source and origin:				
	Model:				
	Should be compatible with the system and the new technology offered				
	Wide field of view not less than 140 deg.				
	Observation range should be between 3-100 mm or better				
	Bending capability: Up / Down (200 / 90 deg or wider) and Right / Left (100 / 100 deg or wider)				
	Insertion tube diameter not more than 10 mm				
	Distal end diameter not more than 10 mm				
	Working channel diameter 3.2 mm				
	Working length not less than 1000 mm				
	Include Water Jet function				
	Waterproof and one-step connector				
	Complete with all accessories and consumables needed to work perfectly				
# 5	HD Colonoscope, Qty: 4				
	Manufacturer:				
	Source and origin:				
	Model:				
	Should be compatible with the system and the new technology offered				



	Wide field of view not less than 170 deg.				
	Observation range should be between 5-100 mm or better				
	Bending capability: Up / Down (180 / 180 deg or wider) and Right / Left (160 / 160 deg or wider)				
	Insertion tube diameter not more than 12.8 mm				
	Distal end diameter not more than 13.4 mm				
	Working channel diameter 3.8 mm				
	Working length not less than 1680 mm				
	Include Water Jet function				
	Waterproof and one-step connector				
	Complete with all accessories and consumables needed to work perfectly				
# 6	Leakage Tester, Qty: 2				
	Manufacturer:				
	Source and origin:				
	Model:				
	To examin all scopes for leakage				
	Should be made from the same manufacturer of the scopes				
# 7	Online UPS suitable for the station, Qty: 2				
# 8	Endoscopic Suction Pump, Qty: 1				
# 9	Co2 insufflator, Qty: 2				
	Manufacturer:				
	Source and origin:				
	Model:				
	Operating Modes: Low, Medium, and High CO2 Flow Rates				
	With Co2 hose british standard				
	Complete with all accessories and consumables needed to work perfectly				
# 10	Irrigation pump, Qty: 2				
	Manufacturer:				
	Source and origin:				

	Model:				
	Maximum Flow Rate: 1) Instrument channel: 730 ml/min 2) Water channel: 220 ml/min Or better				
	With Foot switch, pump head and Water container				
	Complete with all accessories and consumables needed to work perfectly				
# 11	Recorder, DVD, endoscopy Optional to be priced seperately				
	Manufacturer:				
	Source and origin:				
	Model:				
	System for the recording (capture and storage) of video-endoscopy procedures and images in digital media, (DVD / SD / USB etc., combination to be proposed)				
	Suitable for the preparation of medical records				
	Standard common video and images format types input compatibility				
	Internal memory storage capacity				
	Possibility of control via scope head buttons or footswitch; alternative input tools should be described				
	User interface with controls and display				
	Optional to be priced separately:				
	LED module				
	Suction valve				
	Water valve				
	Safty cabinet for hold 4 scops				
	Complete with all accessories and consumables needed to work completely as specified				
	Basic service training for two hospital engineers included in the offer				



Your offer should contain the following:

- 1) Preventive maintenance schedual
- 2) Refernce list for the same model
- 3) Compliance sheet
- 4) Cataloges
- 5) Data sheet
- 6) Autherization latter
- 7) CE or FDA certificate

Warranty: 5 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations

Should the company offer its service way, preferable having back up units and response time should mentioned in the offer

يرجى الإلتزام بتعبئة الجدول التالي :

	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

lm



طرح عطاء مدفوع

يعين مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11429 والخاص بشراء Complete Flexible Endoscopy Station على
الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض اسماؤها في موعد أقصاه يوم الخميس الموافق 2021/10/14 الساعة الثالثة بعد الظهر 03:00 pm وفقا
للاتي:

#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	موعد الطرح	وقت الإغلاق	
					التاريخ	الساعة
1	RFQ-H-11429	Complete Flexible Endoscopy Endoscopy Station	500 شيكل غير مستقرة	04-10-2021	14-10-2021	

حسب الشروط التالية:

1. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الاعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 500 شيكل غير مستقرة .
4. كغالة دخول عطاء بقيمة (2000) النا دولار امريكي .

على الراغبين بالاشتراك في العطاء المذكور طلبه بواسطة ارسال رسالة على الايميل التالي tender3@najah.edu او مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 \ في مستشفى النجاح الوطني الجامعي مبنى الإدارة ،
للحصول على نسخة من وثائق العطاء .

للاستفسار يرجى الاتصال على:

دائرة اللوازم والمشتريات في مستشفى النجاح الوطني الجامعي - البريد الالكتروني tender3@najah.edu هاتف +970-09-2331471 فاكس: +970-9-2389685 .