

التاريخ : 26/05/2022 م

طرح عطاء رقم – RFQ-H-11498

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : **Complete Laparoscopy System K4**، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (250) شيكل إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة - الطابق الأول .
3. كفالة دخول عطاء بقيمة (1000) ألف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم التأمين لاينظر في عرض سعركم .
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر .
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
7. ضرورة أن تكون الأجهزة المعروضة جديدة وليست مجددة .
8. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها .
9. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر وتحملكم فروق الأسعار .
10. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي .
13. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
14. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
15. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
16. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
17. يرجى الالتزام بتعبئة كافة وثائق العطاء من الصفحة 1 ولغاية الصفحة 6 .
18. آخر موعد لتقديم عرض السعر هو يوم الإثنين الموافق 2022-06-06 الساعة الثالثة بعد الظهر PM03:00 .
19. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات
إياد المكاوي



طرح عطاء رقم – RFQ-H-11498

جدول الأسعار والكميات :

No	NNUH Code	Item	QTY	Unit	CODE #	دعوى / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDD01310	Complete Laparoscopy System K4 حسب المواصفات التالية :	1	Pcs					
-		Total Price (NIS)							

#	Complete Laparoscopy system 4K, Qty.:1	Compliance	Deviation
	Good brand name with FDA or CE preferable (Eurob, USA , Japan)		
	for General Surgery, Urology		
	Consist of :		
# 1	TROLLEY		
	Manufacturer:		
	Source and origin:		
	Model:		
	Portable on 4 antistatic dual castors, 2 with locking brakes		
	Required number of shelves for housing all the units of the set		
	Adjustable arm for fixation to either side for fixing the monitor		
	One drawer unit with lock and key		
	Cable Manager		
# 2	MONITOR, Qty.: 2		
	Manufacturer:		
	Source and origin:		
	Model:		
	HDTV Display in 16: 10 HDTV format.		

	one of 32" and one of 26" Medical grade Full HD, display compatable with 4K		
	SDI/HD-SDI, Composite, S-Video, RGB, DVI-D and VGA input		
	All required cables and connectors, which should be specified		
	Dustproof and Drip water protected		
# 3	4 K CAMERA CONTROL UNIT & CAMERA		
	Manufacturer:		
	Source and origin:		
	Model:		
	Corresponds to the 4K		
	Progressive scan technology both on camera head and console		
	The system should have zooming facility to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used.		
	System should be able to optimize all the settings and should be ready as soon as connected to camera control unit with automatic brightness control.		
	Should be compatible for remote controlled operation of various features		
	Video Outputs Composite to RGB, 3G-SDI, DVI, HDMI		
	Camera settings (e.g. white balance, zoom, gain, sharpness etc.) should be possible directly from the camera head buttons.		
	integratable to PACS system, Connected to documentation system, if you have any solution please specify?		
# 4	LIGHT SOURCE:		
	Manufacturer:		
	Source and origin:		
	Model:		
	LED Tybe		
	Manual and automatic adjustment of light intensity		
	Adaptors for karl storz, wolf and olympus		
	LED Lamp life, Please specify?		



	Standby mode		
	Power supply voltage: 220 VAC		
# 5	SUCTION and IRRIGATION PUMP		
	Manufacturer:		
	Source and origin:		
	Model:		
	Fully automated irrigation pump for urology (roller pump)		
	suction and irrigation volume of up to 2 l / min		
	Controlled suction and irrigation unit		
	Irrigation pressure control between 0-450 mm Hg		
	Suction pressure control between 0.6 and 0.75 bar		
	Control from control panel and/or foot pedal		
	Main unit with digital display		
	Overflow protection on suction bottles		
	Accessories should include silicone suction tubings set , reusable pressure domes, bacterial filter and suction bottles with cap		
# 6	INSUFFLATOR		
	Manufacturer:		
	Source and origin:		
	Model:		
	Fully automatic, electronically controlled gas fill		
	45L high flow insufflation unit		
	heater, reusable circuit, adult ant pediatric		
	Real time pressure (RTP) reusable circuit		
	Optical and acoustic warning signals in case of malfunction or excessive pressure.		
	Connectible to medical gas pipeline with filter		
	contain display mode provides clear visualization of the pressure, flow rate and volume in real time for monitoring		
	automatic smoke evacuation		
	CO2 switch-over valve accessory, the surgeon can freely switch CO2 tanks without interrupting long procedures, thus reducing the physical difficulty of exchanging tanks		

lm 

# 7	Telescopes (of 4K quality)		
	1-Telescope 5 mm - 30 degree angle of view 30-33 cm long		
	2-Telescope 5 mm - 0 degree angle of view 30- 33 cm long		
	3-Telescope 10 mm - 30 degree angle of view 30-35 cm long		
	4-Telescope 10 mm - 0 degree angle of view 30-35 cm long		
	All Telescopes should be:		
	Low risk of object burn		
	Colour coded for identification		
	Autoclavable		
	Fibreoptic light transmission incorporated		
	Optional to be priced separately:		
	LED lamp for light source		
	Telescope 5 mm - 30 degree angle of view 30- 33 cm long		
	Telescope 5 mm - 0 degree angle of view 30-33 cm long		
	Telescope 10 mm - 30 degree angle of view 30- 35 cm long		
	Telescope 10 mm - 0 degree angle of view 30-35 cm long		
	Complete with all accessories and consumables needed to work completely as specified		
	Your offer should contain the following: 1) Preventive maintenance schedule 2) Reference list for the same model 3) Compliance sheet 4) Catalogs 5) Data sheet 6) Authorization letter 7) CE or FDA certificate		

Warranty: 3 years, with commitment to
conduct preventive maintenance (labor and
parts) during the warranty period and according
to manufacturer recommendations



Finish



يرجى الإلتزام بتعبئة الجدول التالي :

	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6