



طرح عطاء

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11509 والنخاص بأجهزة طبية ضمن الشروط والمواصفات الموضحة في كراسة وثائق العطاء

وقت الإغلاق		موضوع العطاء	رقم العطاء	#
الساعة	التاريخ			
P.M 03:00	2022/07/27	أجهزة طبية	RFQ-H-11509	1

حسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الإعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 750 شيكل غير مستردة .
4. كفاية دخول عطاء بقيمة (3000) ثلاثة آلاف دولار أمريكي .

على الراغبين بالحصول على نسخة العطاء مراسلتنا على الإيميل التالي tender3@najah.edu او مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 \ في مستشفى النجاح الوطني الجامعي مبنى الإدارة ، للحصول على نسخة من وثائق العطاء .

لاي استفسار بإمكانكم مراجعة :

دائرة اللوازم والمشتريات في مستشفى النجاح الوطني الجامعي – البريد الالكتروني tender3@najah.edu هاتف +970-09-2331471 فاكس: +970-9-2389685 .

التاريخ : 14/07/2022 م

طرح عطاء رقم – RFQ-H-11509

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء – المحترمين .

الموضوع : أجهزة طبية ، حسب المواصفات المرفقة .

ترغب مستشفى النجاح الوطني الجامعي في شراء أجهزة طبية وضمن المواصفات المذكورة في الجداول المرفقة وترحب بالشركات الراغبة والمؤهلة للمشاركة في العطاء وضمن الشروط التالية :

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة, في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق, فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (750) شيكل إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة – الطابق الأول .
3. كفالة دخول عطاء بقيمة (3000) ثلاثة آلاف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأمينا نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم التأمين لاينظر في عرض سعركم.
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأمينا نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر تستبدل بكفالة صيانة بقيمة 5% من قيمة الإحالة ولمدة تغطي مدة الكفالة المجانية .
5. رسوم الاعلان على من يرسو عليه العطاء .
6. على كل مشارك بحال عليه العطاء كاملاً أو جزء من العطاء بمبلغ يزيد عن 2500 شيكل تقديم شهادة خصم مصدر ، وفي حالة عدم تقديمها سيتم خصم 10% من قيمة البنود المحالة لصالح وزارة المالية .
7. يشترط في الأصناف الموردة أن تكون غير مستعملة وغير مجددة أو خضعت للصيانة أو أن تكون معروضة سابقاً ، وأن تكون في صناديق المصنع الأصلية وظروف التخزين لا تؤثر على هيئة أو صلاحية تلك الصناديق أو محتوياتها .
8. **ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .**
9. في حال تأخير التوريد ما يزيد عن اسبوعين يصبح للمستشفى الحق بفرض غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم ، ويحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر وتحملكم فروق الأسعار دون تحمل أي مسؤولية تذكر.
10. الدفع خلال **120** يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة اشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي **بالظرف المختوم فقط** وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.
13. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
14. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
15. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .

16. **اخر موعد لتقديم عرض السعر هو يوم الأربعاء الموافق 2022-07-27 الساعة الثالثة بعد الظهر 00:03 PM .**

17. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه , يرجى الاتصال بالعهد مدير دائرة اللوازم والمشتريات هاتف رقم (97092389687+) داخلي (6400) أو من خلال البريد الإلكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



طرح عطاء رقم – RFQ-H-11509

جدول الأسعار والكميات :

No.	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDD01326	Femtosecond laser machine	1	Pcs					
2	G-A-MDD00259	Phaco vitrectomy machine .	1	Pcs					
3	G-A-MDD00064	Excimer laser machine	1	Pcs					
-		Total Price (NIS)							

جدول المواصفات الفنية :

Item1 #	Femtosecond laser machine, Qty: 1	Compliance	Deviation
	Good brand name with FDA or CE		
	Manufacturer:		
	Source and origin:		
	Model:		
	Femtosecond laser machine Specification:		
	Laser type: femtosecond laser		
	state of the art technology and the latest version from the manufacturer		
	Intelligence eye tracking		
	Pulse duration 220–580 fs		
	Wavelength 1043 nm		
	Laser pulse rate more than 400 kHz		
	Integrated interio chamber OCT		
	With high end corneal topography device		
	Integrated uninterruptible power supply (UPS)		
	power supply: 220-240 V AC; 50 Hz		
	Motorised swivelling patient bed		
	Treatment type:		
	Flap, lenticule extraction, CIRCLE, ICR, Keratoplasty and presbyopia treatment and Astigmatic keratectomy (AK, RK)		

should be upgradable for femtosecond assisted cataract surgery		
Please specify which treatment type need licenses		
Describe any accessories you recommend for us?		
Patient interface for single use, 5 kit		
Adjustable surgical microscope with Magnification: 0.7 X		
multi doctor user account		
With Surgeon's chair		
Service training for one hospital engineer at manufacturer site		
Operation training for one doctor		
Optional to be priced separately:		
Patient interface for single use		
femtosecond assisted cataract surgery software		
Digital video camera for recording surgical procedures		
Should has CB test and EMC test (60601.1 and 60601.2)		
Complete with all accessories and consumables needed to work completely as specified		
Your offer should contain the following: 1) Preventive maintenance schedule 2) Reference list for the same model 3) Compliance sheet 4) Catalogues 5) Data sheet 6) Authorization letter 7) CE or FDA certificate 8) The service and operational manuals should delivered as soft copy with the units		

Warranty: 5 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.



Item 2#	Phaco / Vitrectomy machine, Qty: 1	Compliance	Deviation
	Good brand name with FDA or CE		
	Manufacturer:		
	Source and origin:		
	Model:		
	Phaco / Vitrectomy machine Specification:		
	1) It should be-micron incision vitrectomy surgery and micron incision cataract surgery compatible.		
	2) It should be have dual linear vitrectomy control.		
	3) It should be have effective viscous fluid injection and extraction mode.		
	4) It should be programmable to store various parameters of 20 guage, 23, 25 guage and 27 guage vitrectomy .		
	5) It should have inbuilt system of phacoemulsification for cataract surgery and fragmatome for dropped lens piece removal.		
	6) It should be have user programmable ultrasound duty cycle.		
	7) It should have user combinations of both air and fluid venting systems.		
	8) It should have multifunctional foot switch with regular and dula linear control.(hand switch)		
	9) Reflux should be available		
	10) Programme modes : Vitrectomy, diathermy, fragmatome, U/S 1. U/S 2. U/S3. I/A1, I/A2, I/A3.		
	11) Display of parameters should be available.		
	12)FOOT PEDAL: Programmable Foot Pedal controls all the equipment functions Wireless via blue tooth technology and aback up wire, single Linear Dual Linear Pedal movement up and dawn.		
	13)Interface: Articulating Tray Arm ,Touch screen, logical and intuitive design, Function always in the same screen location , Simple and easy to read , graphics Files can be updated during or following procedure		
	14)USP backup (Software) should be provided		
	Anterior:		
	Phaco technique :		

Standard coaxial (3.2,3.0,2.8,2.65)		
Coaxial small incision 2.2 mm		
Biaxial microincision 1.4 mm		
Coaxial microincision 2.8 mm		
Phacoemulsification (U/S):		
Continuous ultrasound		
Pulsed ultrasound		
Single burst ultrasound		
Multiple burst ultrasound		
Continuous burst ultrasound		
programmable Emission Mode.		
Aspiration:		
Aspiration pump Type: dual pump (Peristaltic and Venturi)		
User can switch between the two pumps during surgery		
Flow rate range aspiration: 1-50 cc/min		
Control Panel or Linear control by the footswitch		
Irrigation:		
Irrigating pressure: Regulated by the height of the built-in electric I.V. pole		
Irrigation control Available by the touch screen or Foot pedal		
Coagulation:		
Type: Bipolar diathermy		
Probes: Forceps and pencil surface probes, endo-ocular pencil probes		
Control by Panel or by the foot pedal		
Operating parameters:		
Type : Liner or fixed		
Range : 1% to 100%		
Anterior vitrectomy:		
Cut rate : 0- 800 cut/min		
Illumination: Self-illuminated cutter		
Posterior:		
Cutter:		
Cutting rate range: Cut rate 0 - 7500 cut/min .		
Cutters Gage: Gage 20,23,25,27		
Control: Panel or linear cut rate control by the foot pedal		
Illumination:		

SOURCES : Xenon or LED light		
Adjustable Light intensity		
What kind of filters do you have : explain if available		
Fiber optic pipes: 20G, 23G, 25G and 27G models		
Scissor:		
Microscissors type: explain if available		
Cutting rate range: From 50 to 300 cuts/min		
Control: Panel or linear cut rate control by the foot pedal		
Sterile air Infusion:		
Auto Air Exchange Valve INCLUDED		
Auto Gas Fill INCLUDED		
Viscous Fluids infusion:		
Auto Fluid Exchange Valve: INCLUDED		
Pressure range: From 0.2 to 5.0 bar (step of 0.2bar)		
Controls: Linear injection control		
Injection system included		
Integrated Laser:		
Probes Straight, curved, illuminated, aspirating,		
Gage: 20-gauge, 23-gauge, 25-gauge, and 27-gauge		
Control multifunctional foot pedal		
Power supply: 220-240 V AC; 50 Hz		
The cassette should have two separate aspiration tubes for both extrusion and cutting		
Accessories to be included:		
Cassettes with tubing system, 3 boxes		
Ultrasound handpieces, Qty.: 2		
Irrigation handpieces, Qty.: 2 and Aspiration handpiece, Qty.: 2		
Anterior vitrectomy cutter, Qty.: 2		
Posterior vitrectomy cutter, Qty.: 2		
Laser probe, Qty.: 2		
Cutary probe, Qty.: 2		
Bimanual irrigation aspiration cannula, 2 boxes		
Phaco tips, 2 boxes		
10 Sleeves & 2-Flare		
Test Chambers, Qty.: 2		
Optional to be priced separately:		
Cassettes with tubing system		
Ultrasound handpieces		

(I/A) handpieces		
Anterior vitrectomy cutter		
Posterior vitrectomy cutter		
Laser probe		
Cutary probe		
Bimanual irrigation aspiration cannula		
Phaco tips		
Sleeves & Flare		
Test Chambers		
Service training for one hospital engineer at manufacturing site		
Should has CB test and EMC test (60601.1 and 60601.2)		
Complete with all accessories and consumables needed to work completely as specified		
Your offer should contain the following: 1) Preventive maintenance schedule 2) Reference list for the same model 3) Compliance sheet 4) Catalogues 5) Data sheet 6) Authorization letter 7) CE or FDA certificate 8) The service and operational manuals should delivered as soft copy with the units		

Warranty: 5 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.



Item3 #	Excimer Laser machine, Qty: 1	Compliance	Deviation
	Good brand name with FDA or CE		
	Manufacturer:		
	Source and origin:		
	Model:		
	Excimer Laser machine Specification:		
	Types: ArF excimer laser		
	state of the art technology and the latest version from the manufacturer		
	wavelength: 193 nm		
	Frequency (pulse rate): 1050 Hz		
	Weight of patient support system: 220 kg or higher		
	power supply: 220-240 V AC; 50 Hz		
	Eye tracking feature with sampling rate 1050 Hz		
	Eye fixation LED		
	The system should automatically centers the ablation of LOAs on this axis.		
	Ablation time should less than 1.5s		
	Motorised swivelling patient bed		
	The system should have adjustable monitor with touch screen		
	The laser head life time, please specify		
	Adjustable surgical microscope		
	Treatment type: LASIK, PRK, Programmable PTK shaping, Topography-guided treatment , PRESBYOPIA TREATMENT non touch PRK (TPRK)		
	Please specify which treatment type need licenses		
	Describe any accessories you recommend for us?		
	multi doctor user account		
	Service training for one hospital engineer at manufacturer site		
	Operation training for one doctor		
	Optional to be priced separately:		
	HD video camera for microscope		
	Should have CB test and EMC test (60601.1 and 60601.2)		

Complete with all accessories and consumables needed to work completely as specified		
Your offer should contain the following: 1) Preventive maintenance schedule 2) Reference list for the same model 3) Compliance sheet 4) Catalogues 5) Data sheet 6) Authorization letter 7) CE or FDA certificate 8) The service and operational manuals should delivered as soft copy with the units		

Warranty: 5 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations and including gas refill

	<u>إسم الشركة</u>	<u>1</u>
	<u>الختم</u>	<u>2</u>
	<u>التاريخ</u>	<u>3</u>
	<u>التوقيع والإسم</u>	<u>4</u>
	<u>معلومات الإتصال</u>	<u>5</u>
	<u>مدة التوريد</u>	<u>6</u>

