

التاريخ: 2022/07/21 م

**إعادة طرح عطاء رقم RFQ-H-11507**

السادة / الشركات الراغبة بالمشاركة في العطاء - المحترمين.

الموضوع: **Anesthesia Machine for MRI**، حسب المواصفات.

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية:

1. تكتب الاسعار بالشيكل الإسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد.
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
3. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، ويحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن اسبوعين.
4. تحديد فترة التسليم في عرض سعركم وأيضاً فترة الضمان، وأن تكون صلاحية العرض المقدم من طرفكم لا تقل عن ثلاثة أشهر.
5. **تسلم عروض الاسعار عبر الإيميل [nnuh.tender@najah.edu](mailto:nnuh.tender@najah.edu) فقط ولن يتم قبول أي عرض سعر إلا من خلاله**، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني.
6. ضرورة الرد على العطاء سلباً أو ايجاباً.
7. ضرورة الالتزام بالتسعير على جدول المواصفات الكميات في نسخة طرح العطاء.
8. الدفع خلال 90 يوم من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة.
9. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار.
10. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، لا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل.
11. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر.
12. اخر موعد لتقديم عرض السعر هو يوم الثلاثاء الموافق 2022/07/26 الساعة العاشرة صباحاً (10:00) AM.
13. لأية استفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (0097092389687) داخلي (6400) أو من خلال البريد الالكتروني [tender3@najah.edu](mailto:tender3@najah.edu).

مع فائق الاحترام ،،،،

مدير دائرة اللوازم والمشتريات

إياد مكاوي



التاريخ: 2022/07/21 م

**إعادة طرح عطاء رقم RFQ-H-11507**

جدول المواصفات والكميات

No.	Descriptions	QTY	Unit	Unit Price (NIS)	Total Price (NIS)
1	Anesthesia Machine for MRI	1	Pcs		
<b>Anesthesia machine for MRI, Qty: 1</b>				<b>Compliance</b>	<b>Deviation</b>
Good brand name with FDA or CE					
Manufacturer:					
Source and origin:					
Model:					
<b>Anesthesia machine for MRI Specification:</b>					
MRI compatible anesthesia machine					
All Patient type: neonatal, pediatric and adult					
Latest safety alarms as applicable to MRI room and gas failure alarms.					
Features facilitate use in both the MRI suite and surgical OR					
Anesthesia machine with - Two Drawers - Top shelf and work area - Extra shelf for patient monitor - Built in ventilator - Central wheels lock - Ventilator screen - Absorber bag arm - Canister - Two vaporizer (Isoflurane and sevoflurane) - CO2 absorber - Flowmeters for Oxygen, air and nitrous oxide - Battery backup - Anesthetic gas scavenging system - Patient circuit and test lugs for adult and pediatrics.					
The Machine should have a built-in anesthesia ventilator					
Should be compact, ergonomic & easy to use with automatic pre-use check for electronic parts.					
Central brake effectively enables fast and easy interlock or release for enhanced safety especially in magnetic fields					
Should be able to deliver a tidal volume from 20ml to 1500ml.					

Anesthesia machine for MRI, Qty: 1	Compliance	Deviation
Should have a battery backup for 60 minutes with low battery alarm and over charge protection.		
Should have monitoring facility of continuous airway pressure, tidal volume, frequency , oxygen concentration and oxygen supply pressure		
High resolution TFT color display (7")		
Should have automatic self test for the entire system.		
Anesthesia machine should be with 3 gas supply system (O2, N2O and Air) with pipeline connections and reserve cylinder yokes.		
Gas cylinder (pin indexed) yokes		
One Pin index yoke for connecting cylinder each for O2, N2O through pipeline		
Should have pressure gauge for all gas inlets		
Oxygen and Nitrous oxide should be linked either mechanically or pneumatically to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture.		
Having mechanical hypoxic guard with automatic cutoff of N2O		
Should provide with oxygen flush switch		
Breathing Frequency (rate) 4 to 60 bpm		
Positive End Expiratory Pressure (PEEP) 0 to 20 cmH2O (hPa)		
Inspiration/expiration ratio (Ti:Te) 4:1 to 1:4		
<b>Ventilation Modes:</b>		
Man/Spon and CPAP		
Volume Controlled (VC)		
Pressure Controlled (PC)		
Pressure Support		
APRV		
SIMV		
<b>The offer includes the following:</b>		
Should provide with adult and pediatric reusable and autoclavable tubing breathing circuit, Qty.1 each		
Isoflurane & sevoflurane vaporizer, Qty.1 each		
O2, Air and N2O hose with british standard		
Two vaporizer (Isoflurane and sevoflurane) with its filling accessories		
Service training for one hospital engineer in the manufacturer site		
<b>Optional to be priced seperately</b>		
System battery pack		
O2 cell		
PM kit		
Gas analyzer		



Anesthesia machine for MRI, Qty: 1	Compliance	Deviation
Your offer should contain the following: 1) Preventive maintenance schedule 2) Reference list for the same model 3) Compliance sheet 4) Catalogues 5) Data sheet 6) Authorization letter 7) CE or FDA certificate 8) The service and operational manuals should be delivered as soft copy with the units		
Warranty: 3 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		

- يرجى من حضرتكم الالتزام بتعبئة المعلومات في الجدول أدناه:

	اسم الشركة والختم	1
	الاسم والتوقيع	3
	معلومات الاتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

LM



إعلان إعادة طرح عطاء MRI

يعلن مستشفى النجاح الوطني الجامعي عن إعادة طرح عطاء رقم RFQ-H-11507، على الشركات المتخصصة والرغبة بالمشاركة بالعطاء تقديم عروض أسعارها في موعد أقصاه يوم الثلاثاء الموافق 2022/07/26 الساعة العاشرة صباحاً AM 10:00 وفقاً للآتي:

وقت الإغلاق		مؤد الطرح	شحن نسخة العطاء	موضوع العطاء	رقم العطاء	#
الساعة	التاريخ					
AM 10:00	26/07/2022	21/07/2022	مجانياً	Anesthesia Machine for MRI	RFQ-H-11507	1

حسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر.

على الراغبين بالاستزاد في العطاء المتكور طلبه بواسطة ارسال رسالة على الايميل التالي [tender3@najah.edu](mailto:tender3@najah.edu) او مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 \ في مستشفى النجاح الوطني الجامعي مبنى الإدارة، للحصول على نسخة من وثائق العطاء.

للاستفسار يرجى الاتصال على:

دائرة اللوازم والمشتريات في مستشفى النجاح الوطني الجامعي - البريد الالكتروني [tender3@najah.edu](mailto:tender3@najah.edu) هاتف +970-09-2331471 فاكس: +970-9-2389685.