

التاريخ : 15/01/2023 م

طرح عطاء رقم – RFQ-H-11563

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .


الموضوع : **Mobile X-Ray** ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (250) شيكل إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة - الطابق الأول .
3. كفالة دخول عطاء بقيمة (1000) ألف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم التأمين لاينظر في عرض سعركم.
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر.
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
7. ضرورة أن تكون الأجهزة المقدمة جديدة وليست مجددة .
8. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
9. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر وتحملكم فروق الأسعار.
10. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.
13. ضرورة تحديد تكلفة عقد الصيانة السنوية للسنة التي تلي السنة الأولى المجانية ولمدة 5 سنوات .
14. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
15. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
16. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
17. آخر موعد لتقديم عرض السعر هو يوم الثلاثاء الموافق 2023-01-31م الساعة الثالثة مساءً 03:00 PM .
18. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



طرح عطاء رقم – RFQ-H-11563

جدول الأسعار والكميات :

No	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDD00177	Mobile X-Ray حسب المواصفات الفنية	1	Pcs					
-	Total Price (NIS)								

يرجى الإلتزام بتعبئة الجدول التالي :

	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

المواصفات الفنية :-

Digital mobile X-ray machine, Qty: 1	Compliance	Deviation
Good brand name with FDA or CE (Europe, USA, Japan)		
Should has CB test and EMC test (60601.1 and 60601.2)		
Manufacturer:		
Source and origin:		
Model:		
Digital Mobile X-ray machine Specification:		
State of the art technology and the latest version from the manufacturer and the latest and best technology on the market		
The unit must have an effective braking system for parking, transport and emergency braking.		
The tube stand must be fully counterbalanced with rotation in all directions.		
Exposures with remote control and with corded switch should be possible. Remote control should be offered with system.		
The Dose Area Product (DAP) meter should measure the X-ray dose output at the collimator, with dose report		
SID Meter		
Shoud has flexible telescopic tube arm and sliding column		
The column move up, down and rotation it can move the tub arm longitudinal		
Anti-collision sensor and brake		
With Hook for lead apron and Storage trays		
Please specify the stand by duration?		
Should enable the exposure during charging		
Can use the device if battery empty and during charging		
Power for chargings should be single phase, 220V, 50 Hz AC Supply		
Motorized drive and compact wheelbase		
The device has ability to choose the patient type		
Generator:		
Power, high frequency: 40 kW or higher, if you have higher please offer it as an option		
mA range: 10 to 400 mA		
Exposure times: 0.001 - 10 s		
Focal spot: 1.5		
Tube head movement:		
Vertical movement: around 550-2020 mm		
Tube Extension: around 700-1250 mm		
Column rotation range: around $\pm 300^\circ$		
Active tube head brakes		
Tube overload protection should be available		

Batteries:		
Dual batteries, Separate one for drive and one generator control		
Typical usage 4 hours of continuous movement		
Please specify the duration requested for battery charging?		
Collimator:		
Manual and motorized, with LED light field indicator		
Should has touch screen controller		
Rotation: ±90°		
Motorized additional filtration		
Detector:		
One wireless Detector with two batteries		
Water proof detector		
Shoud the detector charger integrated with the X-ray device		
Detector sizes: large 43 cm x 43 cm		
Image matrix size: 2330 x 2846 pixel		
Detector pixels: 6.5 Megapixels		
Image resolution: up to 3.30 Lp/mm		
Maximum patient weight: 250 Kg		
With detector cover or handle		
Connectivity, interface and memory:		
The machine should be fully network ready and it should be possible to transfer images and patient data from and to hospital network using LAN connectivity and wireless LAN.		
System should be DICOM compatibilit		
It should provide the possibility to write all Patient images, Studies and single images onto CDs directly on workstation.		
Should has monitor 17 inch touch-screen monitor		
Image Storage more than 3500 images		
Training included in the price:		
Service training for one hospital engineer at manufacturer site and must be real service training		
Optional to be priced separately:		
1- Detector battery		
2- Small detector 25 cm X 25 cm for pediatric		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		

[Handwritten signature]

Complete with all accessories and consumables needed to work completely as specified		
--	--	--

Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		

Warranty: 5 years, including the software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations, including all parts including the tube.		
--	--	--





طرح عطاء مدفوع

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11563 والخاص بـ Mobile X-Ray ، على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض اسعارها في موعد اقصاه يوم الثلاثاء الموافق 31/01/2023 الساعة الثالثة عصراً 03:00 PM :

وقت الإغلاق	التاريخ	موعد الطرح	ثمن نسخة العطاء	موضوع العطاء	رقم العطاء	#
P.M 03:00	2023/01/31م	2023/01/15م	250 شيكل غير مستردة	Mobile X-Ray	RFQ-H-11563	1

حسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الإعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 250 شيكل غير مستردة .
4. كفاية دخول عطاء بقيمة (1000) ألف دولار أمريكي .

على الراغبين بالحصول على نسخة العطاء مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 / مبنى الإدارة أو عبر الإيميل Tender3@najah.edu وذلك خلال ساعات الدوام الرسمية .

للاستفسار يرجى الاتصال على :

لأية إستفسارات أخرى يرجى مراجعة دائرة اللوازم والمشتريات هاتف رقم 970-09- 2331471- +970-9-2389685 : فاكس: 970-9-2389685 .